

PACKAGE

Tester Only



Tester with Stand and Foot Plate



Tester with Stainless Steel Foot Plate



CUSTOM TEST LIMIT INSTRUCTIONS

Use the forms below to select the high and low test limits for the footwear and wrist strap test circuits. Fail points are determined by the factory. They are typically 20% higher and lower than the pass limit.

*The Combo Tester must have the same low test limit for both the footwear and wrist strap circuits. The low test limit is shared by both circuits.

FOOTWEAR TEST LIMITS

High Limit Pass: Default Limit - 100 megohms (1.0×10^8 ohms)

Custom Limit (500 megohms max):

***Low Limit Pass:** Default Limit - 750 kilohms (7.5×10^5 ohms)

Custom Limit (5 kilohms min):

WRIST STRAP TEST LIMITS

High Limit Pass: Default Limit - 10 megohms (1.0×10^7 ohms)

Custom Limit (150 megohms max):

***Low Limit Pass:** Default Limit - 750 kilohms (7.5×10^5 ohms)

Custom Limit (5 kilohms min):

Company Name: _____
Address: _____
City/State: _____
Country: _____ ZIP Code: _____
Contact Name: _____
Phone: _____
Email: _____
Signature: _____

Special manufactured items are not returnable for credit.

Internal use only:

Quote ID: _____ Person Taking Order: _____

Item No: _____ Work Order No: _____

DESCO CUSTOM COMBO TESTER WORKSHEET

DESCO

DESCO WEST: 3651 WALNUT AVE, CHINO, CA 91710 WEBSITE: Desco.com
PHONE: (909) 627-8178

DESCO EAST: ONE COLGATE WAY, CANTON, MA 02021
PHONE: (781) 821-8370

Desco Custom Combo
Tester Worksheet

DATE:
December
2017